

## EC DECLARATION OF CONFORMITY

In accordance with EN ISO 17050-1:2004

We Fluke Precision Measurement Ltd.

T/A Comark Instruments

Of Hurricane way

Norwich, NR6 6JB United Kingdom

In accordance with the following directives:

EC Directive 2004/108/EC Electromagnetic Compatibility (EMC)

EC Directive 2011/65/EU RoHS

hereby declare that:

Equipment: K Type Thermometer

Model number: C28

is in conformity with the applicable requirements of the following documents:

EMC

EN61326-1:2006

<u>Safety</u>

EN61010-1:2010 (Third Edition)

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The equipment complies with all applicable Essential Requirements of the Directive

Signed by:

Name: David Cooper

Position: Site Director and Engineering Manager

For: Fluke Precision Measurement Ltd. T/A Comark Instruments

On: 12 Nov 14